


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>1643</b>  Accredited to <b>ISO/IEC 17025:2017</b>	<b>Cavendish Laboratories Limited</b>	
	Issue No: 034    Issue date: 01 February 2022	
	<b>Millers Barn</b> The Warren Estate Lordship Road Writtle Chelmsford Essex CM1 3WT	<b>Contact: Mr P M Jarvis</b> Tel: +44 (0)1245 422800 E-Mail: <a href="mailto:info@cavendishlaboratories.com">info@cavendishlaboratories.com</a> Website: <a href="http://www.cavendishlaboratories.com">www.cavendishlaboratories.com</a>
Testing performed by the Organisation at the locations specified below		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> Millers Barn The Warren Estate Lordship Road Writtle Chelmsford Essex CM1 3WT	<b>Local contact</b> Mr P M Jarvis  Tel: +44 (0)1245 422800 Email: <a href="mailto:info@cavendishlaboratories.com">info@cavendishlaboratories.com</a> Website: <a href="http://cavendishlaboratories.com">cavendishlaboratories.com</a>	Health and Hygiene Head Office  Asbestos – All Support Functions  A

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Testing Laboratories	Health and Hygiene	C



1643  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Cavendish Laboratories Limited**  
**Issue No: 034 Issue date: 01 February 2022**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	B, C
	Sampling of air for fibre counting	Documented In-House Method in Instruction Manual APP A1 Section 17: P5B based on HSG 248	
	Fibre counting	Documented In-House Method Part 3A, Section1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 stage clearance process	Documented In-House Method Part 3A, Section1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B, C
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method Part 3B Sections 1-3 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A,
MICRO ORGANISMS IN AIR	Sampling of air for TVC and Yeasts and Moulds	Documented In-House Method P5A using RCS Air Sampler	B
Air Samples	<u>Microbiological Tests</u>		
	Isolation and enumeration:	Documented In-House Methods	
	Total Viable Count	P5AS2	A
ENVIRONMENTAL SWABS	Yeasts and Moulds	P5AS2	A
	<u>Microbiological Tests</u>	Documented In-House Methods	
	Total Viable Count	P3CS4	A
	Yeasts and Moulds	P3CS4	A



1643  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Cavendish Laboratories Limited**  
**Issue No: 034 Issue date: 01 February 2022**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS	<u>Microbiological Tests</u>		
Potable	Isolation and enumeration:  Coliforms <i>Escherichia coli</i>	Documented In-House Methods based on "The Microbiology of Drinking Water" 2002:  P3CS2 – 2.4 using Idexx MPN method based on "The Microbiology of Drinking Water (Part 4)" 2016	A
	Total Viable Count @ 22°C and 37°C	P3C3S2 – 2.3 Pour Plate Method based on "The Microbiology of Drinking Water (Part 7)" 2020	A
	Confirmed <i>Pseudomonas aeruginosa</i>	P3CS2 – 2.5 using membrane filtration based on "The Microbiology of Drinking Water (Part 8)" 2015	A
Industrial process waters, cooling towers, other waters, showers, spa pools, swimming pools and sediments	<i>Legionella pneumophila</i> serogroups 1, 2-14 and <i>Legionella spp</i>	Documented In-House Method No P3CS3 – 3.0 based on BS 6068-4.12:1998 (withdrawn)	A
	Coliforms <i>Escherichia coli</i>	P3CS2 – 2.4 using Idexx MPN method based on "The Microbiology of Drinking Water (Part 4)" 2016	A
	Total Viable Count @ 22°C, 30°C and 37°C	P3C3S2 – 2.3 Pour Plate Method based on "The Microbiology of Drinking Water (Part 7)" 2020	A
	Confirmed <i>Pseudomonas aeruginosa</i>	P3CS2 – 2.5 using membrane filtration based on "The Microbiology of Drinking Water (Part 8)" 2015	A
END			