


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 1643 Accredited to ISO/IEC 17025:2017	Cavendish Laboratories Limited Issue No: 035 Issue date: 10 June 2022	
	Millers Barn The Warren Estate Lordship Road Writtle Chelmsford Essex CM1 3WT	Contact: Rachel Knight Tel: +44 (0)1245 422800 E-Mail: info@cavendishlaboratories.com Website: www.cavendishlaboratories.com

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Millers Barn The Warren Estate Lordship Road Writtle Chelmsford Essex CM1 3WT Local contact Rachel Knight Tel: +44 (0)1245 422800 Email: info@cavendishlaboratories.com Website: cavendishlaboratories.com	Health and Hygiene Head Office Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Testing Laboratories	Health and Hygiene	C



1643
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Cavendish Laboratories Limited
Issue No: 035 Issue date: 10 June 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method in Instruction Manual APP A1 Section 17: P5B based on HSG 248	B
	Fibre counting	Documented In-House Method Part 3A, Section1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 stage clearance process	Documented In-House Method Part 3A, Section1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method Part 3B Sections 1-3 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
ENVIRONMENTAL SWABS	<u>Microbiological Tests</u>	Documented In-House Methods	
	Total Viable Count	P3CS4	A
	Yeasts and Moulds	P3CS4	A



1643
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Cavendish Laboratories Limited
Issue No: 035 Issue date: 10 June 2022

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS	<u>Microbiological Tests</u>		
Potable	Isolation and enumeration: Coliforms <i>Escherichia coli</i>	Documented In-House Methods based on "The Microbiology of Drinking Water" 2002: P3CS2 – 2.4 using Idexx MPN method based on "The Microbiology of Drinking Water (Part 4)" 2016	A
	Total Viable Count @ 22°C and 37°C	P3C3S2 – 2.3 Pour Plate Method based on "The Microbiology of Drinking Water (Part 7)" 2020	A
	Confirmed <i>Pseudomonas aeruginosa</i>	P3CS2 – 2.5 using membrane filtration based on "The Microbiology of Drinking Water (Part 8)" 2015	A
Industrial process waters, cooling towers, other waters, showers, spa pools, swimming pools and sediments	<i>Legionella pneumophila</i> serogroups 1, 2-14 and <i>Legionella spp</i>	Documented In-House Method No P3CS3 – 3.0 based on BS 6068-4.12:1998 (withdrawn)	A
	Coliforms <i>Escherichia coli</i>	P3CS2 – 2.4 using Idexx MPN method based on "The Microbiology of Drinking Water (Part 4)" 2016	A
	Total Viable Count @ 22°C, 30°C and 37°C	P3C3S2 – 2.3 Pour Plate Method based on "The Microbiology of Drinking Water (Part 7)" 2020	A
	Confirmed <i>Pseudomonas aeruginosa</i>	P3CS2 – 2.5 using membrane filtration based on "The Microbiology of Drinking Water (Part 8)" 2015	A
END			